

ORGENTEC Diagnostika GmbH

Carl-Zeiss-Straße 49
55129 Mainz - Germany

Phone: +496131/92580
Fax: +496131/925858
Internet : www.orgentec.com



Instruction for use
August 2005



ORG 520 Anti-Insulin

Immunometric Enzyme Immunoassay for the quantitative determination of autoantibodies to bovine, porcine and recombinant human insulin

CONTENTS

NAME AND INTENDED USE
SUMMARY AND EXPLANATION OF THE TEST
PRINCIPLE OF THE TEST
WARNINGS AND PRECAUTIONS
CONTENTS OF THE KIT
STORAGE AND STABILITY
MATERIALS REQUIRED
SPECIMEN COLLECTION AND HANDLING
PROCEDURAL NOTES
PREPARATION OF REAGENTS
TEST PROCEDURE
INTERPRETATION OF RESULTS
PERFORMANCE CHARACTERISTICS
LIMITATIONS OF PROCEDURE
INTERFERING SUBSTANCES
REFERENCES
INCUBATION SCHEME

NAME AND INTENDED USE

Anti-Insulin is an indirect solid phase enzyme immunoassay (ELISA) for the quantitative measurement of IgG class autoantibodies against bovine, porcine and recombinant human insulin in human serum or plasma. The assay is intended for in vitro diagnostic use only as an aid in the diagnosis of type I diabetes.

SUMMARY AND EXPLANATION OF THE TEST

Type I Diabetes is mainly characterised by limited or fully missing secretion of the hormone insulin. Morphological studies demonstrated a destruction of the beta cells of the so-called Langerhans'sche Cells (Islet Cells) in Type I diabetics. Numerous researchers described the appearance of antibodies directed against the islet cells and insulin as the causal reason for the onset of the disease.

Anti-Insulin antibodies are found in 37 percent of patients with newly detected Type I Diabetes, in 4 percent of their relatives of the first degree and in up to 1,5 percent of healthy controls. A positive correlation between the appearance of anti-Insulin and anti-Islet Cell antibodies has been reported.

Anti-Insulin autoantibodies may be detected several months and in some cases years before the onset of the fully clinical manifestation of the diseases. Occasionally also autoantibodies to Pro-Insulin may appear.

These "true" anti-Insulin autoantibodies directed against endogenous insulin have to be distinguished from those autoantibodies which are developed in insulin dependent diabetics undergoing therapy with insulin preparations of animal origin. In fact the latter have to be referred to side effects. These side effects may occur as local reactions of the skin by development of insulin-specific autoantibodies. These autoantibodies are causing the formation of an insulin depot and they may simulate a resistance against the hormonal treatment with animal insulin.

Additionally other immunological phenomenon have been reported for Type I diabetics. A lot of other autoantibody specificities have been detected in those patients too, but these antibodies must not cause additional autoimmune phenomenon.

The table below shows the incidence (in percent) of various autoantibodies occurring in Type I diabetics compared to healthy controls:

Autoantibody specificity:	Type I diabetics (%)	healthy controls (%)
Anti-Islet Cell antibodies	32	1
Antibodies against Islet Cell surface antigens	approx. 50	approx. 2
Anti-Insulin antibodies	up to 70	0
Anti-Thyroid peroxidase ab's (Anti-TPO)	18	6
Anti-single stranded DNA ab's (Anti-ssDNA)	85	9

Indications:

- anti-Insulin antibodies in Type I diabetics
- development of anti-Insulin antibodies under insulin therapy

PRINCIPLE OF THE TEST

A mixture of highly purified preparations of bovine, porcine and recombinant human insulin is bound to microwells. Antibodies against these antigens, if present in diluted serum or plasma, bind to the respective antigen. Washing of the microwells removes unbound serum and plasma components. Horseradish peroxidase (HRP) conjugated anti-human IgG immunologically detects the bound patient antibodies forming a conjugate/antibody/antigen complex. Washing of the microwells removes unbound conjugate. An enzyme substrate in the presence of bound conjugate hydrolyzes to form a blue color. The addition of an acid stops the reaction forming a yellow end-product. The intensity of this yellow color is measured photometrically at 450 nm. The amount of colour is directly proportional to the concentration of IgG antibodies present in the original sample.

WARNINGS AND PRECAUTIONS

1. All reagents of this kit are strictly intended for in vitro diagnostic use only.
2. Do not interchange kit components from different lots.
3. Components containing human serum were tested and found negative for HBsAg, HCV, HIV1 and HIV2 by FDA approved methods. No test can guarantee the absence of HBsAg, HCV, HIV1 or HIV2, and so all human serum based reagents in this kit must be handled as though capable of transmitting infection.
4. Avoid contact with the TMB (3,3',5,5'-Tetramethyl-benzidine). If TMB comes into contact with skin, wash thoroughly with water and soap.
5. Avoid contact with the Stop Solution which contains hydrochloric acid (1 M). If it comes into contact with skin, wash thoroughly with water and seek medical attention.
6. Some kit components (i.e. Controls, Sample buffer and Buffered Wash Solution) contain Sodium Azide as preservative. Sodium Azide (NaN_3) is highly toxic and reactive in pure form. At the product concentrations (0.09%), though not hazardous. Despite the classification as non-hazardous, we strongly recommend using prudent laboratory practices (see 8., 9., 10.)
7. Some kit components contain Proclin 300 as preservative. When disposing reagents containing Proclin 300, flush drains with copious amounts of water to dilute the components below active levels.
8. Wear disposable gloves while handling specimens or kit reagents and wash hands thoroughly afterwards.
9. Do not pipette by mouth.
10. Do not eat, drink, smoke or apply makeup in areas where specimens or kit reagents are handled.
11. Avoid contact between the buffered Peroxide Solution and easily oxidized materials; extreme temperature may initiate spontaneous combustion.

Observe the guidelines for performing quality control in medical laboratories by assaying controls and/or pooled sera. During handling of all kit reagents, controls and serum samples observe the existing legal regulations.

CONTENTS OF THE KIT

Package size	96 determ.
Qty.1	Divisible microplate consisting of 12 modules of 8 wells each, coated with mixture of highly purified preparations of bovine, porcine and recombinant human insulin. Ready to use.
6 vials, 1.5 ml each	Calibrators with IgG class Anti-Insulin antibodies (A-F) in a serum/buffer matrix (PBS, BSA, NaN_3 <0,1% (w/w)) containing: 0; 6.3; 12.5; 25; 50; and 100 U/ml. Ready to use.
2 vials, 1,5 ml each	Anti-Insulin Controls in a serum/buffer matrix (PBS, BSA, NaN_3 <0,1% (w/w)) positive (1) and negative (2), for the respective concentrations see the enclosed package insert. Ready to use.
1 vial, 20 ml	Sample buffer (Tris, NaN_3 <0,1% (w/w)), yellow, concentrate (5x).
1 vial, 15 ml	Enzyme conjugate solution (PBS, PROCLIN 300 <0,5% (v/v)), (light red) containing polyclonal rabbit anti-human IgG; labelled with horseradish peroxidase. Ready to use.
1 vial, 15 ml	TMB substrate solution. Ready to use.
1 vial, 15 ml	Stop solution (1 M hydrochloric acid). Ready to use.
1 vial, 20 ml	Wash solution (PBS, NaN_3 <0,1% (w/w)), concentrate (50x).

STORAGE AND STABILITY

1. Store the kit at 2-8 °C.
2. Keep microplate wells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light during storage and usage.
5. Diluted sample buffer and wash buffer are stable for at least 30 days when stored at 2-8 °C.

MATERIALS REQUIRED

Equipment

- Microplate reader capable of endpoint measurements at 450 nm
- Multi-Channel Dispenser or repeatable pipet for 100 μl
- Vortex mixer
- Pipets for 10 μl , 100 μl and 1000 μl
- Laboratory timing device
- Data reduction software

Preparation of reagents

- Distilled or deionized water
- Graduated cylinder for 100 and 1000 ml
- Plastic container for storage of the wash solution

SPECIMEN COLLECTION, STORAGE AND HANDLING

1. Collect whole blood specimens using acceptable medical techniques to avoid hemolysis.
2. Allow blood to clot and separate the serum by centrifugation.
3. Test serum should be clear and non-hemolyzed. Contamination by hemolysis or lipemia is best avoided, but does not interfere with this assay.
4. Specimens may be refrigerated at 2-8 °C for up to five days or stored at -20 °C up to six months.
5. Avoid repetitive freezing and thawing of serum samples. This may result in variable loss of autoantibody activity.
6. Testing of heat-inactivated sera is not recommended.

PROCEDURAL NOTES

1. Do not use kit components beyond their expiration dates
2. Do not interchange kit components from different lots
3. All materials must be at room temperature (20-28 °C)
4. Have all reagents and samples ready before start of the assay. Once started, the test must be performed without interruption to get the most reliable and consistent results.
5. Perform the assay steps only in the order indicated
6. Always use fresh sample dilutions
7. Pipette all reagents and samples into the bottom of the wells
8. To avoid carryover contamination change the tip between samples and different kit controls
9. It is important to wash microwells thoroughly and remove the last droplets of wash buffer to achieve best results.
10. All incubation steps must be accurately timed.
11. Control sera or pools should routinely be assayed as unknowns to check performance of the reagents and the assay.
12. Do not re-use microplate wells.

For all controls, the respective concentrations are provided on the labels of each vial. Using these concentrations a calibration curve may be calculated to read off the patient results semi-quantitatively.

PREPARATION OF REAGENTS

Preparation of sample buffer

Dilute the contents of each vial of the sample buffer concentrate (5x) with distilled or deionized water to a final volume of 100 ml prior to use. Store refrigerated: stable at 2-8 °C for at least 30 days after preparation or until the expiration date printed on the label.

Preparation of wash solution

Dilute the contents of each vial of the buffered wash solution concentrate (50x) with distilled or deionized water to a final volume of 1000 ml prior to use. Store refrigerated: stable at 2-8 °C for at least 30 days after preparation or until the expiration date printed on the label.

Sample preparation

Dilute all patient samples 1:100 with sample buffer before assay. Therefore combine 10 µl of sample with 990 µl of sample buffer in a polystyrene tube. Mix well. Controls are ready to use and need not be diluted.

TEST PROCEDURE

1. Prepare a sufficient number of microplate modules to accommodate controls and prediluted patient samples.
2. Pipet **100 µl** of calibrators, controls and prediluted patient samples in duplicate into the wells.

	1	2	3	4	5	6
A	SA	SE	P1	P5		
B	SA	SE	P1	P5		
C	SB	SF	P2	P..		
D	SB	SF	P2	P..		
E	SC	C1	P3			
F	SC	C1	P3			
G	SD	C2	P4			
H	SD	C2	P4			

SA-SF: standards A to F
P1, P2...C: patient sample 1, 2 ...
C1: positive control
C2: negative control

3. Incubate for 30 minutes at room temperature (20-28 °C)
4. Discard the contents of the microwells and wash 3 times with **300 µl** of wash solution.
5. Dispense **100 µl** of enzyme conjugate into each well
6. Incubate for 15 minutes at room temperature
7. Discard the contents of the microwells and wash 3 times with **300 µl** of wash solution
8. Dispense **100 µl** of TMB substrate solution into each well
9. Incubate for 15 minutes at room temperature
10. Add **100 µl** of stop solution to each well of the modules and incubate for 5 minutes at room temperature
11. Read the optical density at 450 nm and calculate the results. Bi-chromatic measurement with a reference at 600-690 nm is recommended.

The developed colour is stable for at least 30 minutes. Read optical densities during this time.

Automation

The ORGENTEC Anti-Insulin ELISA is suitable for use on open automated ELISA processors. The test procedure detailed above is appropriate for use with or without automation.

INTERPRETATION OF RESULTS

Quality Control

This test is only valid if the optical density at 450 nm for Positive Control (1) and Negative Control (2) as well as for the Calibrator A and F complies with the respective range indicated on the Quality Control Certificate enclosed to each test kit! If any of these criteria is not fulfilled, the results are invalid and the test should be repeated.

Calculation of results

For Anti-Insulin a 4-Parameter-Fit with lin-log coordinates for optical density and concentration is the data reduction method of choice. Smoothed Spline Approximation and log-log coordinates are also suitable.

Recommended Lin-Log Plot

First calculate the averaged optical densities for each calibrator well. Use lin-log graph paper and plot the averaged optical density of each calibrator versus the concentration. Draw the best fitting curve approximating the path of all calibrator points. The calibrator points may also be connected with straight line segments. The concentration of unknowns may then be estimated from the calibration curve by interpolation.

Calculation example

The figures below show typical results for Anti-Insulin ELISA. These data are intended for illustration only and should not be used to calculate results from another run.

Calibrators									
No	Position	OD 1	OD 2	Mean	Conc. 1	Conc. 2	Mean	decl. Conc.	CV %
STA	A 1/B 1	0.036	0.032	0.034	0.1	0.1	0.1	0.0	8
STB	C 1/D 1	0.354	0.345	0.349	6.3	6.1	6.2	6.3	2
STC	E 1/F 1	0.621	0.602	0.611	12.9	12.4	12.7	12.5	2
STD	G 1/H 1	0.984	1.005	0.994	25	25	25	25	1
STE	A 2/B 2	1.503	1.500	1.502	50	50	50	50	0
STF	C 2/D 2	2.057	2.035	2.046	102	99	100	100	1

Interpretation of results

In a normal range study with serum samples from healthy blood donors the following ranges have been established with the anti-Insulin test:

	Anti-Insulin [U/ml]
normal:	< 10
elevated:	> 10

Positive results should be verified concerning the entire clinical status of the patient. Also every decision for therapy should be taken individually. It is recommended that each laboratory establishes its own normal and pathological ranges. The reference ranges below should be regarded as guidelines only.

PERFORMANCE CHARACTERISTICS

Parallelism

In dilution experiments sera with high antibody concentrations were diluted with sample buffer and assayed in the Anti-Insulin kit.

Sample	Dilution	Observed [U/ml]	Expected [U/ml]	O/E
1	1:50	77.6		
	1:100	41.7	38.8	107 %
	1:200	21.1	19.4	109 %
	1:400	10.3	9.7	106 %
	1:800	4.7	4.9	96 %
2	1:100	100.7		
	1:200	50.7	50.4	101 %
	1:400	23.7	25.2	94 %
	1:800	11.1	12.6	88 %
	1:1600	5.3	6.3	84 %

Precision (Reproducibility)

Statistics for coefficients of variation (CV) were calculated for each of three samples from the results of 24 determinations in a single run for Intra-Assay precision. Run-to-run precision was calculated from the results of 5 different runs with 6 determinations of each sample:

Intra-Assay			Inter-Assay		
Sample No	Mean [U/ml]	CV [%]	Sample No	Mean [U/ml]	CV [%]
1	11.2	2.5	1	11.6	6.0
2	27.6	2.9	2	31.2	5.2
3	59.7	4.0	3	69.5	4.3

Sensitivity

The lower detection limit for anti-Insulin has been determined at 0.5 U/ml.

Specificity

The microplate is coated with a mixture of highly purified preparations of bovine, porcine and recombinant human insulin. Therefore the anti-Insulin test kit recognises only IgG-class autoantibodies specific for these insulins. No crossreactivities have been observed.

Calibration

Since no international reference preparation for anti-Insulin autoantibodies is available, the assay system is calibrated in relative arbitrary units.

LIMITATIONS OF PROCEDURE

The Anti-Insulin ELISA is a diagnostic aid. A definite clinical diagnosis should not be based on the results of a single test, but should be made by the physician after all clinical and laboratory findings have been evaluated.

INTERFERING SUBSTANCES

No interference has been observed with haemolytic (up to 1000 mg/dL), lipemic (up to 3 g/dL triglycerides) or bilirubin (up to 40 mg/dL) containing sera. Nor have any interfering effects been observed with the use of anticoagulants. However for practical reasons it is recommended that grossly hemolyzed or lipemic samples should be avoided.

REFERENCES

1. Reeves, W.G. Insulin Antibody Determination: Theoretical and Practical Considerations. *Diabetologia* 24, 399 (1983).
2. Atkinson, M.A., Fisk, D.D., Spillar, R.P. and MacLaren, N.K. Insulin antibodies as markers for Insulin dependent diabetes mellitus (IDD). *Diabetes* 34, 926 - 930 (1985).
3. Willein, T., Nicholson, S. and Casey, C. A Micro Enzyme-Linked Immunosorbent Assay for Insulin Antibodies in Serum. *J. Imm. Methods* 76, 185 (1985).
4. Kobayashi, N. et al. A Solid-Phase Enzyme Immunoassay for Anti-Insulin Antibody in Diabetes Mellitus Patients. *J. Imm. Methods* 84, 245 (1985).
5. Wissllein, T. et al. Value of Insulin Antibodies as Serum marker for Insulin-Dependent Diabetes Mellitus. *Lancet* 1, 480 (1985).
6. Soeldner, J.S. et al. Insulin-Dependent Diabetes Mellitus and Autoimmunity: islet-cell Autoantibodies and beta-cell failure. *New Engl. J. Med.* 313, 893 (1985).
7. Torfs, C.P. et al. Long Term Frozen Sera for Epidemiological Studies of Antibodies. *Lancet* 1, 503 (1986).
8. Seino, S. et al. Characterisation of Circulating Insulin in Insulin autoimmune Syndrome. *Clin. Endo. & Metab.* 62, 64 (1986).
9. Boitard, C.H. and McDewitt, H.O. Immunology of insulin-dependent diabetes mellitus. In: Cohen, J.R., (ed.) *Perspectives on autoimmunity*, 39-58 (1987). CRC Press, Boca Raton, FL
10. Nouvo, J.A., Baker, Jr., J.R., Wartowsky, I. et al. Autoantibodies to insulin are present in sera of patients with autoimmune thyroid diseases. *Diabetes*, 37, 317-320 (1988).
11. Witkin, T.J. Insulin Autoantibodies as markers for type I diabetes. *Endocrine Reviews*, 11, 92-104 (1990).

INCUBATION SCHEME

